

INTRODUCTION

Serology is central to the screening and diagnosis of syphilis infection and can be classified into two groups:

1. Non-treponemal tests: Venereal Disease Research Laboratory (VDRL) or Rapid Plasma Reagin (RPR)
2. Treponemal tests which are based on antigens derived from *Treponema pallidum*: Treponemal Haemagglutination Assay (TPHA), Fluorescent Treponemal Antibody absorption test (FTA-Abs) and native or recombinant antigen Enzyme immunoassay (EIA) tests.

AIM

In this study, the performance of 5 automated tests, using recombinant proteins (Table I), was compared to that of TPHA (*Treponema pallidum* hemagglutination assay, Biokit, Barcelona, Spain):

- Architect Syphilis TP (Abbott, Delkenheim, Germany);
- BioPlex® 2200 Syphilis IgG, IgM (BioRad, Bio-Rad, Marnes la Coquette, France),
- Elecsys® Syphilis (Roche, Diagnostics GmbH, Mannheim, Germany),
- IMMULITE® 2000 Syphilis (Siemens, Healthcare Diagnostics, Llanberis, United Kingdom),
- LIAISON® Treponema Screen (DiaSorin, Saluggia, Italy).

Table I: Recombinant proteins used in these 5 automated assays.

| Assays | Support automate | Antigen Recombinant | Ab detection | Run time (min) |
|---------------------------|------------------|-----------------------|--------------|----------------|
| Architect Syphilis TP | Architect | TpN15, TpN17 TpN47 | Total Ab | 26 |
| BioPlex® Syphilis IgG | BioPlex 2200 | TpN15, TpN17 TpN47 | IgG | 45 |
| BioPlex® Syphilis IgM | BioPlex 2200 | TpN17, TpN47 | IgM | 45 |
| Elecsys® Syphilis | Cobas | TpN15, TpN17 TpN47 | Total Ab | 18 |
| IMMULITE® Syphilis | IMMULITE 2000 | TpN17 | Total Ab | 35 |
| LIAISON® Treponema Screen | LIAISON | TpN17 | Total Ab | 39 |

METHODS

392 unselected sera (except Elecsys®: 385 unselected sera, due to availability of limited volume of samples) and 7 selected samples from patients with primary syphilis were tested by TPHA as well as these 5 assays. Whenever one assay scored reactive, IgG and IgM Line Immunoblot (Genzyme Virotech GmbH, Rüsselsheim, Germany) was used for confirmation.

RESULTS

The following tables summarize the results obtained for the 392 unselected samples tested with each method (equivocal results were excluded).

Concordance between TPHA and five automated assays

Table II: Architect vs TPHA

| Architect Syphilis TP | TPHA | | | Total |
|-----------------------|----------|-----------|----------|-------|
| | Positive | Equivocal | Negative | |
| Positive | 259 | 0 | 6* | 265 |
| Equivocal | 0 | 0 | 1 | 1 |
| Negative | 7** | 0 | 119 | 126 |
| Total | 266 | 0 | 126 | 392 |

*3/6 of the discordant samples were positive by Immunoblot IgG
 **0/7 of the discordant samples were positive by Immunoblot IgG

Percent Agreement

| | | |
|----------|-------|-----------|
| Positive | 97.4% | (259/266) |
| Negative | 95.2% | (119/125) |
| Overall | 96.7% | (378/391) |

Table III: BioPlex® 2200 Syphilis IgG vs TPHA

| BioPlex® 2200 Syphilis IgG | TPHA | | | Total |
|----------------------------|----------|-----------|----------|-------|
| | Positive | Equivocal | Negative | |
| Positive | 259 | 0 | 7* | 266 |
| Equivocal | 1 | 0 | 1 | 2 |
| Negative | 7** | 0 | 117 | 124 |
| Total | 267 | 0 | 125 | 392 |

*3/7 of the discordant samples were positive by Immunoblot IgG
 **0/7 of the discordant samples were positive by Immunoblot IgG

Percent Agreement

| | | |
|----------|-------|-----------|
| Positive | 97.4% | (259/266) |
| Negative | 94.3% | (117/124) |
| Overall | 96.4% | (376/390) |

Table IV: Elecsys® vs TPHA

| Elecsys® Syphilis | TPHA | | | Total |
|-------------------|----------|-----------|----------|-------|
| | Positive | Equivocal | Negative | |
| Positive | 256 | 0 | 4* | 260 |
| Equivocal | 0 | 0 | 0 | 0 |
| Negative | 5** | 0 | 120 | 125 |
| Total | 261 | 0 | 124 | 385 |

*2/4 of the discordant samples were positive by Immunoblot IgG

**0/5 of the discordant samples were positive by Immunoblot IgG

Percent Agreement

| | | |
|----------|-------|-----------|
| Positive | 98.1% | (256/261) |
| Negative | 96.8% | (120/124) |
| Overall | 97.7% | (376/385) |

Table V: IMMULITE® vs TPHA

| IMMULITE® 2000 Syphilis Screen | TPHA | | | Total |
|--------------------------------|----------|-----------|----------|-------|
| | Positive | Equivocal | Negative | |
| Positive | 260 | 0 | 6* | 266 |
| Equivocal | 0 | 0 | 0 | 0 |
| Negative | 6** | 0 | 120 | 126 |
| Total | 266 | 0 | 126 | 392 |

*3/6 of the discordant samples were positive by Immunoblot IgG

**0/6 of the discordant samples were positive by Immunoblot IgG

Percent Agreement

| | | |
|----------|-------|-----------|
| Positive | 97.7% | (260/266) |
| Negative | 95.2% | (120/126) |
| Overall | 96.9% | (380/392) |

Table VI: LIAISON® Treponema Screen vs TPHA

| LIAISON® Treponema Screen | TPHA | | | Total |
|---------------------------|----------|-----------|----------|-------|
| | Positive | Equivocal | Negative | |
| Positive | 259 | 0 | 8* | 267 |
| Equivocal | 0 | 0 | 1 | 1 |
| Negative | 7** | 0 | 117 | 124 |
| Total | 266 | 0 | 126 | 392 |

*3/8 of the discordant samples were positive by Immunoblot IgG

**0/7 of the discordant samples were positive by Immunoblot IgG

Percent Agreement

| | | |
|----------|-------|-----------|
| Positive | 97.4% | (259/266) |
| Negative | 93.6% | (117/125) |
| Overall | 96.2% | (376/391) |

Architect, BioPlex® (IgG), Elecsys®, IMMULITE® and LIAISON® show versus TPHA an overall agreement of 96.7%, 96.4%, 97.7%, 96.9% and 96.2% respectively. Their specificity is greater than that of TPHA, based on the Immunoblot results obtained for the discordant samples. Also their sensitivity is greater, since they were able to detect three (LIAISON®, IMMULITE®, BioPlex®, Architect) and two (Elecsys®) more samples as positive compared to TPHA.

Concordance between LIAISON® and four other automated assays

The LIAISON® Treponema test shows an overall agreement of 99.5%, 99.7%, 99.2% and 99.7% versus Architect Syphilis TP, BioPlex® 2200 Syphilis (IgG), IMMULITE® 2000 Syphilis Screen and Elecsys® Syphilis respectively (table VIII).

Table VII: Correlation between results of LIAISON® and four other automated assays

| LIAISON® | Architect | | | BioPlex® IgG | | | Elecsys® | | | IMMULITE® | | |
|-----------|-----------|--------|------|--------------|--------|------|----------|--------|------|-----------|--------|------|
| | Pos. | Equiv. | Neg. | Pos. | Equiv. | Neg. | Pos. | Equiv. | Neg. | Pos. | Equiv. | Neg. |
| Positive | 267 | 0 | 2 | 265 | 2 | 1 | 260 | 0 | 1 | 267 | 0 | 2 |
| Equivocal | 0 | 0 | 0 | 1 | 0 | 1 | 0 | 0 | 1 | 0 | 0 | 1 |
| Negative | 0 | 0 | 122 | 0 | 0 | 122 | 0 | 0 | 123 | 1 | 0 | 121 |

Percent Agreement

| | | | | | | | | |
|----------|-------|-----------|-------|-----------|-------|-----------|-------|-----------|
| Positive | 100% | (267/267) | 100% | (265/265) | 100% | (260/260) | 99.6% | (267/268) |
| Negative | 98.4% | (122/124) | 99.2% | (122/123) | 98.4% | (123/124) | 98.4% | (121/123) |
| Overall | 99.5% | (389/391) | 99.7% | (387/388) | 99.7% | (383/384) | 99.2% | (388/391) |

In addition, 7 primary syphilis specimens (i.e., TPHA-negative) scored positive by Architect, BioPlex® IgM, Elecsys®, IMMULITE® and LIAISON® (Table VIII). By contrast, two patients scored negative by BioPlex® IgG.

Table VIII: Primary Syphilis samples

| | VDRL | TPHA | Architect | Elecsys® | IMMULITE® | LIAISON® | BioPlex® IgM | BioPlex® IgG | Immunoblot IgM | | | | | |
|---|------|------|-----------|----------|-----------|----------|--------------|--------------|----------------|-----|-----|-----|-----|----|
| | | | Abbott | Roche | Siemens | DiaSorin | TpN | TpN | TpN | TpN | TpN | TpN | TpN | |
| 1 | N | N | P | P | P | P | 17 | 47 | 15 | 17 | 47 | 47 | 17 | 15 |
| 2 | N | N | P | P | P | P | N | P | N | N | N | P | N | P |
| 3 | N | N | P | P | P | P | N | P | N | P | P | P | P | P |
| 4 | P | N | P | P | P | P | P | P | N | P | N | P | P | P |
| 5 | N | N | P | P | P | P | P | P | N | P | P | P | P | N |
| 6 | N | N | P | P | P | P | N | P | N | P | N | P | P | N |
| 7 | P | N | P | P | P | P | P | P | P | P | P | P | P | P |

N: negative P: positive

DISCUSSION

Architect Syphilis TP, BioPlex® 2200 Syphilis, LIAISON® Treponema, IMMULITE® 2000 Syphilis Screen and Elecsys® Syphilis have excellent sensitivity and specificity performance. LIAISON® shows an overall agreement > 99% versus Architect, BioPlex® (IgG), IMMULITE® and Elecsys®. These five fully-automated immunoassays show an overall agreement > 96% versus TPHA and their specificity and sensitivity is greater than that of TPHA, based on the Immunoblot results obtained for the discordant samples. This evaluation with 7 primary syphilis clearly demonstrates the greater sensitivity of these five fully-automated immunoassays to that of VDRL and TPHA.

CONCLUSION

These five fully-automated immunoassays show an overall agreement > 96% versus TPHA. With their suitability for automation Architect, Elecsys®, IMMULITE® and LIAISON® are an ideal screening test allowing correct identification of all 7 primary infections with one test only. In this study, incorporation of only one recombinant *Treponema* antigen (TpN17, LIAISON®) is not associated to a lower sensitivity in comparison to the use of three antigens (TpN15, TpN17, TpN47). The BioPlex® IgM helps to differentiate past infections from recent ones as demonstrated by the negative BioPlex® IgG results for 2 of the 7 primary samples.